

ESSENTIAL TEST DATA

Complies with European Standard EN 14885:2022 and the latest regulatory requirements for biocidal activity of high-level surface disinfectants applied with mechanical action (e.g., wiping).

A $\geq 5 \log_{10}$ reduction was achieved for bacteria, and a $\geq 4 \log_{10}$ reduction for viruses, fungi, yeasts, mycobacteria, and bacterial spores in the following tests. Additionally, the requirements of the 4-field tests EN 16615 and EN 17846 were met, with F2–F4 contamination levels being $< 50 \text{ CFU/cm}^2$.

ORGANISM	TEST METHOD	TEST TYPE	CONDITIONS
SPORICIDAL			
Bacillus subtilis	EN 17126 (P2, S1)	Suspension	Clean 1 & Dirty 1
Bacillus cereus			
Clostridioides difficile			
Clostridioides difficile	EN 17846 (P2, S2)	Surface with mechanical action	Clean 1 & Dirty 1
MYCOBACTERICIDAL			
Mycobacterium terrae	EN 14348 (P2, S1)	Suspension	Clean 1 & Dirty 1
Mycobacterium avium			
VIRUCIDAL			
Poliovirus Type 1	EN 14476 (P2, S1)	Suspension	Clean 1 & Dirty 1
Adenovirus Type 5			
Murine norovirus			
FUNGICIDAL/YEASTICIDAL			
Aspergillus brasiliensis	EN 13624 (P2, S1)	Suspension	Clean 1 & Dirty 1
Candida albicans			
Candida albicans	EN 16615 (P2, S2)	Surface with mechanical action	Clean 1 & Dirty 1
BACTERICIDAL			
Pseudomonas aeruginosa	EN 13727 (P2, S1)	Suspension	Clean 1 & Dirty 1
Staphylococcus aureus			

ORGANISM	TEST METHOD	TEST TYPE	CONDITIONS
BACTERICIDAL			
<i>Enterococcus hirae</i>	EN 13727 (P2, S1)	Suspension	Clean 1 & Dirty 1
<i>Pseudomonas aeruginosa</i>	EN 16615 (P2, S2)	Surface with mechanical action	Clean 1 & Dirty 1
<i>Staphylococcus aureus</i>			
<i>Enterococcus hirae</i>			

ADDITIONAL TEST DATA

Beyond essential testing, further studies have confirmed the product's effectiveness against key pathogens and demonstrated broader performance characteristics.

MICROORGANISMS

ORGANISM	TEST METHOD	TEST TYPE	CONDITIONS
BACTERIAL SPORES			
<i>Bacillus cereus</i>	EN 13704 (P2, S1)	Suspension	Dirty 1
<i>Bacillus subtilis var niger</i>	Babb JR, Bradely CR & Ayliffe GAJ (J. of Hosp. Inf. 1980 1:63-75)	Suspension	Clean 1 & Dirty 3
PROTOZOA			
<i>Acanthamoeba castellanii</i> cysts	Bespoke Test	Suspension	Clean 1
MYCOBACTERIA			
<i>Mycobacterium terrae</i>	EN 14563 (P2, S2)	Carrier	Clean 1 & Dirty 2
<i>Mycobacterium avium</i>			
<i>Mycobacterium terrae</i>	EN 16615 (P2, S2)	Surface with mechanical action	Dirty 1
<i>Mycobacterium avium</i>			
<i>Mycobacterium terrae</i>	Griffiths et al. (J. of Hosp. Inf. 1998 38:183-92)	Suspension	Clean 1 & Dirty 4
VIRUSES			
Adenovirus Type 5	EN 17111 (P2, S2)	Carrier	Clean 1 & Dirty 1

ORGANISM	TEST METHOD	TEST TYPE	CONDITIONS
VIRUSES			
Murine norovirus	EN 17111 (P2, S2)	Carrier	Clean 1 & Dirty 1
Polyomavirus SV40			
Poliovirus Type 1	ASTM E-1053	Surface without mechanical action	Dirty 2
Adenovirus Type 5			
Feline Calicivirus			
Hepatitis B Virus (HBV)			
Herpes Simplex Virus (HSV) Type 1			
Human Immunodeficiency Virus (HIV)			
Influenza A Virus (H1N1)			
Adenovirus Type 5	EN 16615 (P2, S2)	Surface with mechanical action	Dirty 1
Murine norovirus			Clean 1 & Dirty 1
Bovine Coronavirus			Dirty 1
Influenza A Virus (H1N1)	EN 14476 (P2, S1)	Suspension	Dirty 1
Parvovirus (Using Minute Virus of Mice (MVM) surrogate)	DVV/RKI	Suspension	Dirty 3
Poliovirus Type 1			Clean 2 & Dirty 3
Adenovirus Type 5			
Murine norovirus			
Polyomavirus SV40			
Vaccinia Virus			
FUNGI & YEASTS			
Aspergillus brasiliensis	EN 14562 (P2, S2)	Carrier	Clean 1
Candida albicans			Dirty 2
Candidozyma auris (Formerly Candida auris)			

ORGANISM	TEST METHOD	TEST TYPE	CONDITIONS	
FUNGI & YEASTS				
Aspergillus brasiliensis	EN 16615 (P2, S2)	Surface with mechanical action	Clean 1 & Dirty 1	
Candida albicans	AOAC Use-Dilution Method	Carrier	Dirty 2	
Candida albicans	EN 13697 (P2, S2)	Surface without mechanical action	Clean 1	
Fusarium solani	EN 13624 (P2, S1)	Suspension	Clean 1	
Aspergillus sydowii				
BACTERIA				
Pseudomonas aeruginosa	EN 14561 (P2, S2)	Carrier	Clean 1	
Staphylococcus aureus				
Enterococcus hirae				
Carbapenem-resistant Klebsiella pneumoniae (CRE)				
Vancomycin-resistant Enterococcus faecium (VRE)				Dirty 2
Multidrug-resistant Acinetobacter baumannii (MDRAB)				
Extended Spectrum Beta-Lactamase (ESBL) producing Klebsiella pneumoniae				
Methicillin-resistant Staphylococcus aureus (MRSA)				
Pseudomonas aeruginosa	EN 13697 (P2, S2)	Surface without mechanical action	Clean 1	
Staphylococcus aureus				
Enterococcus hirae				
Escherichia coli				
Proteus vulgaris	EN 16615 (P2, S2)	Surface with mechanical action	Dirty 1	
Methicillin-resistant Staphylococcus aureus (MRSA)			Clean 1	
Neisseria gonorrhoeae				
Gardnerella vaginalis				

ORGANISM	TEST METHOD	TEST TYPE	CONDITIONS
BACTERIA			
<i>Streptococcus agalactiae</i>	EN 16615 (P2, S2)	Surface with mechanical action	Clean 1
Methicillin-resistant <i>Staphylococcus aureus</i> (MRSA)	EN 13727 (P2, S1)	Suspension	Dirty 1

BIOFILM

ORGANISM	TEST METHOD	BIOFILM TYPE	SURFACE TYPE
<i>Pseudomonas aeruginosa</i>	MBEC assay (ASTM E2799-22)	Grown in moist conditions - aged for 72 hours	Polystyrene
<i>Staphylococcus aureus</i>	CDC Biofilm Reactor (ASTM E2871-22)		Stainless steel & PVC
<i>Staphylococcus aureus</i>	Modified CDC Biofilm Reactor	Dry (semi-hydrated) - aged for 12 days	Stainless steel & PVC

OTHER

CAPABILITY	TEST METHOD
DNA/RNA Degradation	Polyacrylamide gel electrophoresis (PAGE)

KEY

Clean 1	0.3 g/l Bovine albumin	Dirty 3	10% Blood Serum
Clean 2	Aqua bidest	Dirty 4	1% Blood Serum
Dirty 1	3g/l Bovine albumin + 3ml Blood erythrocytes	P2, S1	Phase 2, Step 1
Dirty 2	5% Blood Serum	P2, S2	Phase 2, Step 2