

Tristel Protect™


Transportation bags for instruments
with integrated traceability




User guide

How to use Tristel Protect


Tristel Protect bags are designed specifically for the transportation of instruments, with traceability in mind. Each bag is identified with a unique and patented label, incorporating a resealable pocket. This pocket enables you to store traceability data, without compromising the integrity of the instrument inside the bag.



Take one Tristel Protect bag and one Tristel Protect label from the box.




Place the clean instrument inside the bag.




Double fold the top of the bag and seal with label as shown.

Note: The intact (solid blue) label confirms that the instrument inside the bag is clean.




Peel the corner of the label and secure traceability data within the marked area. Reseal the pocket.

Note: The instrument is now ready to be transported to the patient. Upon arrival to the patient, remove traceability data from the pocket and attach it to patient records.



Open bag to reveal an "OPEN VOID" marking and remove the instrument. Do not throw the bag away.



Following use, place the dirty instrument inside the bag and reseal it using the label.

Note: The OPEN VOID marking on the label confirms that the instrument inside the bag is dirty. The pocket label can be used again to complete the traceability between device, disinfection procedure and patient. Discard the bag to clinical waste once cycle is complete.

Tristel

Created by: Tristel Solutions Limited, Lynx Business Park, Cambs, UK, CB8 7NY
T +44 (0) 1638 721500 - E mail@tristel.com - W www.tristel.com

Copyright © 2016 Tristel Solutions
TRU/337/1
November 2017