

# Tristel Trio™

## Wipes System



### TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Advanced Tactile Imaging	Endocavity Transducer	No Series Specified	Vaginal Tactile Imager (VTI)
Bard	Ultrasound Unit	No Series Specified	Site-Rite 8
	Skin Surface Transducer	No Series Specified	32MM Linear Probe and 20MM Linear Probe
BK Medical / BK Ultrasound	Skin Surface Transducer	BK3000/BK5000	6C2, 6C2s, 9C2, 14L3, 13L4W, 10L2W, 18L5, 8L2
	Endocavity Transducer		E13C2, E14C4t, E14C4, E14 CL4b, E10C4, 20R3, X14L4
	Intraoperative Transducer		I14C5I, I14C5T, I12C5b, I12C4f
	Skin Surface Transducer		N13C5, N11C5s, X18L5s, X12C4
	Ultrasound Transducer	Flex Focus & Pro Focus	5P1
Carestream	Skin Surface Transducer	Touch Prime	8809, 8814, 8815, 8816, 8824, 8826, 8836, 8862, 8863, 2050, 2052, 8808, 8808e, 8818, 8838, 8848
	Intraoperative Transducer		6C2, 6C2s, 9C2, 5P1, N13C5, 8L2
	Endocavity Transducer		18L5s, 18L5, 14L3, 13L4W, 10L2W
Clarius Mobile Health	Handheld Ultrasound Scanners	No Series Specified	Clarius Scanner
			Arthrex Synergy MSK
Echosens	FibroScan and Ultrasound Unit	FibroScan 520 TOUCH	M6.0, D5-1268, M3.5, D7-1257, M2.5, D10-1269
Esaote	Endocavity Transducer	No Series Specified	EC123*, EC1123*
	Intraoperative Transducer	No Series Specified	IOE323*
	TOE Transducer	No Series Specified	TEE022*, TEE122*
Exact Imaging	Endocavity Transducer	ExactVu	EV9C, EV29L
Fujifilm	Endocavity Transducer	Ultrasonic Mini Probe System	PB-2020 M
GE Healthcare	Ultrasound Transducer	Logiq Systems - E9	C1-5-D, C1-6-D, C2-9-D, M6C-D, IC5-9-D, 9L-D, 11L-D, ML6-15-D, L8-18i-D, 6s-D, 6s-D, RAB2-5-D, RAB4-8-D, RIC5-9-D
		Logiq Systems - S8	C1-5-D, IC5-9-D, 9L-D, ML6-15-D
		Logiq Systems - S7	C1-5-D, 8C, IC5-9-D, 9L-D, 11L-D, ML6-15, L8-18i-D, RAB4-8-D
		Logiq P Series	4C, 5CS, 8C, BE9CS, E8C, E8CS, 8L, 9L, 11L, ML6-15, I739, T739, 3S, 5S, 6S, 7S, 4D3C-L, 4D5C-L, 4DE7C
		Logiq C Series	C2-5-RC, E8CS-RC, L6-12-RC, 3S-RC
		Logiq A Series	3.5CS, 5CS, 8C, BE9CS, E8CS, 8L, 9L, 10L, 11L, 12L, I739, T739, 3S, 5S, 7S
		Logiq E Series	4C-RS, 8C-RS, E8C-RS, 8L-RS, 9L-RS, 12L-RS, i12L-RS, L8-18i-RS, I739-RS, T739-RS, 3S-RS, 6S-RS
		Vivid E9	4C-D, 9L-D, 11L-D, ML6-15-D, 6S-D, 12S-D, 4V-D
		Vivid P Series	4C, 5CS, 8C, E8CS, 8L, 11L, T739, 3S, 6S
		Vivid S Series	4C-RS, 8C-RS, E8C-RS, 8L-RS, 9L-RS, 12L-RS, 3S-RS, 3Sc-RS, 5S-RS, 6S-RS, 7S-RS, 12S-RS, M4S-RS

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
GE Healthcare	Ultrasound Transducer	Vivid e/i/q Series	4C-RS, 8C-RS, E8C-RS, 8L-RS, 9L-RS, 12L-RS, 3S-RS, 3Sc-RS, 5S-RS, 6S-RS, 7S-RS, 12S-RS, M4S-RS
		Voluson Expert Series	AB2-7-D, 4C-D, C1-5-D, C4-8-D, M6C-D, IC5-9-D, 9L-D, 11L-D, ML6-15-D, 3S-D, RAB2-5-D, RAB4-8-D, RIC5-9-D, RIC6-12-D, RRE5-10-D
		Voluson Performance Series	4C-RS, E8C-RS, 12L-RS, 3Sc-RS, RAB2-5-RS, RIC5-9W-RS
		Voluson Signature Series	4C-RS, AB2-7-RS, C1-5-RS, 8C-RS, E8C-RS, 9L-RS, 12L-RS, ML6-15-RS, 3Sc-RS, RAB2-5-RS, RAB4-8-RS, RIC5-9W-RS
	Venue 40 Series	4C-SC, E8CS-SC, 12L-SC, L8-18i-SC, 3S-SC	
	TOE Transducer	No Series Specified	6T, 6T-RS, 6Tc, 6Tc-RS, 6Tv, 9T, 9T-RS, 6VT-D
Hitachi	Ultrasound Transducer	No Series Specified	C22l, C22K, C22P, C22T, C25, C251, C252, C25P, C35, C41, C41B, C41L47RP, C41RP, C41V, C41V1, C42, C42K, C42T, CC41R, CC41R1, CL4416R, L34, L43K, L44, L441, L44K, L44LA1, L46K, L46K1, L51K, L53K, L53K1, L55, L64, S12, S121, S21, S211, S22, S31, S31KP, S42, SML44, VC41V, ASU-1000B-3.5, ASU-1000C-3.5, ASU-1001, ASU-1002, ASU-1003, ASU-1005, ASU-1009, ASU-1010, ASU-1011, ASU-1012, ASU-1013, ASU-35-3, ASU-35CWD-2, ASU-35WL-10, ASU-35WSJ-7.5, ASU-36WL-10, ASU-64B, ASU-65B, ASU-67, UST-2265-2, UST-2266-5, UST-5024N-3.5, UST-5037P-3.5, UST-5039P-3.5, UST-5043P-3.5, UST-5044-3.5, UST-5045P-3.5, UST-5046-3.5, UST-5047-5, UST-52103, UST-52105, UST-52108, UST-52109, UST-52114P, UST-52124, UST-52127, UST-52129, UST-5226D-3.5/2, UST-5236D-2.5/2, UST-5242D-2.5/2, UST-5244-3.5, UST-5255-2.5, UST-5256-3.5, UST-5257-5, UST-5263D-2.5/2, UST-5265-2.5, UST-5266-3.5, UST-5267-5, UST-5268P-5, UST-5274-2.5, UST-5275-3.5, UST-5276-5, UST-5279-3.5, UST-5281-5, UST-5283-2.5, UST-5284-2.5, UST-5285-3.5, UST-5286-2.5, UST-5287-3.5, UST-5292-5, UST-5294-5, UST-5295, UST-5296, UST-5297, UST-5298, UST-5299, UST-533, UST-534, UST-535, UST-536, UST-5410, UST-5410H, UST-5411, UST-5412, UST-5413, UST-5415, UST-5417, UST-5419, UST-547, UST-5512U-7.5, UST-5514DTU-7.5, UST-5515-7.5, UST-5521L-7.5, UST-5523L-7.5, UST-5524-5, UST-5524-7.5, UST-5526L-7.5, UST-5531, UST-5534T-7.5, UST-5539-7.5, UST-5540P-7.5, UST-5542, UST-5543, UST-5545, UST-5546, UST-5548, UST-5551, UST-5561-7.5, UST-556T-7.5, UST-556TU-7.5, UST-567, UST-568, UST-5710-7.5, UST-5711-7.5, UST-5712, UST-5713T, UST-574T-7.5, UST-575-7.5, UST-576T-7.5, UST-577-7.5, UST-579T-7.5, UST-5818-5, UST-5819T-5, UST-657-5, UST-660-7.5, UST-665P-5, UST-667P-5, UST-670P-5, UST-675P, UST-676P, UST-677P, UST-9101-7.5, UST-9102-3.5, UST-9102U-3.5, UST-9103-5, UST-9104-5, UST-9111-5, UST-9112-5, UST-9113P-3.5, UST-9114-3.5, UST-9115-5, UST-9116P-5, UST-9118, UST-9119, UST-9120, UST-9121, UST-9123, UST-9124, UST-9125-7.5, UST-9126, UST-9127, UST-9128, UST-9130, UST-9132l, UST-9132T, UST-9133, UST-9135P, UST-9136, UST-9136U, UST-9137, UST-9145, UST-9146l, UST-9146T, UST-9147, UST-9148, UST-9149, UST-9151, UST-932-5, UST-932-7.5, UST-935N-5, UST-939-3.5, UST-939D-3.5, UST-943-3.5, UST-944B-3.5, UST-945B-5, UST-945BP-5, UST-952-5, UST-953P-5, UST-954-5, UST-957-7.5, UST-961-5, UST-961-7.5, UST-964P-5, UST-968P-5, UST-969-5, UST-970-7.5, UST-972-3.5, UST-974-5, UST-975-7.5, UST-977-5, UST-978-3.5, UST-979-3.5, UST-980-5, UST-980P-5, UST-981-5, UST-981P-5, UST-984-5, UST-984P-5, UST-985-7.5, UST-987-7.5,

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Hitachi	Ultrasound Transducer	No Series Specified	UST-988-2.5, UST-989-3.5, UST-989D-3.5, UST-990-5, UST-992-5, UST-994P-5, UST-995-7.5, UST-996P-5, UST-MC11-8731, EUP-B314, EUP-B512, EUP-B514, EUP-B712, EUP-B715, EUP-C321T, EUP-C511, EUP-C514, EUP-C516, EUP-C524, EUP-C532, EUP-C715, EUP-CV714, EUP-F334, EUP-F531, EUP-L34T, EUP-L52, EUP-L53, EUP-L53L, EUP-L53S, EUP-L54MA, EUP-L73S, EUP-L74M, EUP-L75, EUP-LS85, EUP-O33JA, EUP-O53T, EUP-O54J, EUP-O732T, EUP-S50A, EUP-S52, EUP-S63, EUP-S70, EUP-S72, EUP-S80, EUP-V53W, EUP-V73W, EUP-VV731
Mcube	Bladder Scanner	Cubescan	BioCon-700 and BioCon-900
Medcom	Endocavity Transducer	BiopSee Bi Plane Transrectal Transducer	BIPL7.5/70/128, BIC6.5/10/128
Mindray	Ultrasound Transducer	No Series Specified	C5-1s, C5-1E, C5-1U, C5-2, C5-2s, C5-2E, C6-2E, C6-2Gs, C6-2GE, C7-3E, C11-3E, C11-3s, D7-2, D7-2E, D7-2s, D6-2, D6-2P, L7-3, L7-3s, L7-3E, L9-3U, L11-3U, L11-4, L11-4s, L12-3, L12-3E, L12-3s, L12-4, L12-4s, L14-6, L14-6s, L14-6P, L14-6N, L14-6Ns, L14-6NP, L14-6NE, L14-6WE, L14-6Ws, L14-6WU, LM14-6E, LM14-6s, LM16-4U, P4-2, P4-2s, P4-2E, P4-2NE, P10-4E, P10-4s, SC6-1U, SP5-1U, SP5-1s, SP5-1E, V11-3HU, V11-3U, 3C1, 3C1s, 3C1p, 10L24EA, 35C50EA, 35C50EB, 35C50EA, 35C50P, 65C15EAV, 65C15EA, 75L38EA, 75L38EB, 75L53EA, 75L60EA, 75L38P
	TOE Transducer	No Series Specified	P7-3T, P7-3TE
Philips*	Ultrasound Transducer	iU22	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD15	
		HD11 XE	
		HD9	
		HD7 XE	
		HD7	
		HD6	
		HD5	
	Endocavity Transducer	CX50	C9-3v
		Sparq	
		CX30	
		EPIQ	
		HD15	BP10-5ec
		HD11 XE	
		EPIQ	
		Affinity 50	
		Affinity 70	C9-4v
		EPIQ7	
EPIQ5			
Affinity 50			
Affinity 70			
CX50			

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Endocavity Transducer	EPIQ7	C10-3v
		EPIQ5	
		iU22	
		Affinity 70	
		CX50	
	Skin Surface Transducer	EPIQ7	S5-1
		EPIQ5	
		Affinity70	
		CX50	
		iU22	
		iE33	
		Sparq	
		EPIQ7	S12-4
		EPIQ5	
		Affinity70	
		Affinity50	
		iE33	
		HD5	
	HD15		
	HD11		
	CX50		
	HD11 XE		
	No Series Specified	3D6-2	
	No Series Specified	3D8-4	
	Endocavity Transducer	HD11	3D9-3v
		HD15	
		iU22	
		ClearVue	
		EPIQ	
		Affinity	
iU22		7.5 MHz	
Sparq		C8-4v	
Affinity			
Skin Surface Transducer	CX30	C6-2 Compact	
	Sparq		
	Affinity		
	Sparq	C6-3	

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Skin Surface Transducer	CX30	C8-5 Compact
		CX50	
		Affinity	
		EPIQ	
		iE33/iU22	X3-1
		EPIQ7/iE33	X5-1
		EPIQ7/iE33/iU22	X7-2
		CX50	C5-1
		EPIQ7	
		QPIQ5	
		Affinity 70	
		iU22	
		iE33	
		HD15	
		ClearVue	C5-2
		Lumify	
		Xperius	
		iE33	
		iU22	
		HD15	
		HD11 XE	
	HD11		
	HD9		
	HD7 XE		
	HD7		
	HD6		
	HD5		
	EPIQ	C9-2	
	CX50	C9-4	
	No Series Specified	C9-3io	
	Endocavity Transducer	iE33	C9-4v
		EPIQ7	
		EPIQ5	
Affinity50			
Affinity70			
CX50		C9-4ec	
CX50			
EPIQ		C9-4v Compact	

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Endocavity Transducer	CX50	C9-5ec
	Skin Surface Transducer	EPIQ7	C8-5
		EPIQ5	
		Affinity 70	
		Affinity 50	
		iE33	
		iU22	
		HD7 XE	
		HD11 XE	
		HD9	
		HD11	
		HD15	
		CX30	
		CX50	
		Sparq	
	Endocavity Transducer	EPIQ	C10-4ec
	Skin Surface Transducer	iE33	L8-4
		iU22	
		HD11 XE	
		CX50	
		iE33	L9-3
iU22			
HD15			
HD9			
HD11 XE			
CX50			
No Series Specified		L10-4lap	
CX50		L11-3	
iE33			

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Skin Surface Transducer	HD5	L12-3
		HD6	
		HD7	
		HD11	
		HD15	
		Affinity	
		CX50	
		EPIQ	
		Lumify	
		Xperius	
		ClearVue	
		CX30	L12-4 Compact
		Sparq	
		CX50	
		Affinity	
		Sparq	L12-5
		CX50	L12-5 Compact
		Affinity	
		EPIQ	
		iU22	L17-5
		iU22	L18-5
		Sparq	L15-710
		EPIQ7	
		EPIQ5	
		Affinity 70	
		Affinity 50	
		iE33	
		iU22	
		HD11	
	HD15		
	CX30		
	CX50		
TOE Transducer	iE33	T6207	

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model		
Philips*	TOE Transducer	iE33	D2cwc		
		EPIQ7			
		EPIQ5			
		iU22			
		CX50			
		HD11			
		HD9			
		HD15			
		HD11 XE			
	Ultrasound Transducer	No Series Specified	D5cwc		
	TOE Transducer	No Series Specified	X8-2t, X8-2ti		
		EPIQ7	D2tcd		
		EPIQ5			
		iE33			
		iU22			
		HD11 XE			
		EPIQ7	S8-3		
		EPIQ5			
		Affinity50			
		Affinity70			
		CX30			
		CX50			
		iE33			
		iU22			
		HD11			
		HD11 XE			
		HD15			
		Skin Surface Transducer		iU22	S3-1
				iU22	S4-1
	ClearVue				
	Lumify				
	iU22		S4-2		
CX30	S4-2 Compact				
Sparq					
Affinity					
Affinity	S5-1				
EPIQ					



TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model	
Philips*	Skin Surface Transducer	iU22	S5-2	
		iU22	S8	
		CX30	S8-3 Compact	
		CX50		
		HD5		
		Affinity		
		EPIQ		
	TOE Transducer	No Series Specified		S8-3t
				S6-2mpt
		iE33	S7-2t	
		iU22		
		HD15		
		HD11 XE		
		HD11		
		HD9		
		HD7 XE		
		HD7		
		HD6		
		HD5		
		iE33		S7-2 Omni (21778A)
		iU22		
		HD15		
		HD11 XE		
		HD11		
		HD9		
		HD7 XE		
		HD7		
		HD6		
		HD5		
		No Series Specified	S7-3t	
		HD5	S12	
		Skin Surface Transducer	Affinity	S12-4 Compact
			EPIQ	
TOE Transducer		iE33	T6H/21378A OMNI III	
	HD11 XE			
	HD7 XE			

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model	
Philips*	Skin Surface Transducer	HD11 XE	V6-2	
		HD15		
		HD5		
		iU22		
		Affinity		
		ClearVue		
		EPIQ		
		iU22		V8-4
		iU22		VL13-5
		EPIQ		X6-1
	Pencil Transducer	No Series Specified		1.9MHz CW
				5.0MHz CW
				5.0MHz CW Pencil
	TOE Transducer		EPIQ7	X7-2t
			EPIQ5	
Affinity70				
iE33				
CX50				
Sparq				
Samsung*	Ultrasound Transducer	No Series Specified	C2-8ED	
	Skin Surface Transducer		CA1-7A, C2-6IC, CA2-8A, CF2-8, CA2-8AD, L5-13IS, L7-16IS, LA3-16AI, L3-12A, LA3-14AD, L5-13/50, CV1-8A, CV1-8AD, 3DC-6EN, CA2-9A, CA2-9AD, CA1-7AD, CA2-6BM, L3-8ED, LS6-15, LA3-14AD, LA3-14A, PA1-5A, PA1-5AD, SP3-8, CA3-10A	
	Endocavity Transducer		ER4-9/10ED, EV4-9/10ED, V5-9, 3D4-9ES, EVN4-9/10ED, 3D4-9ES, EV3-10B, EVN4-9, EC4-9CD, VR5-9, E3-12A, EA2-11B	
Siemens*	Skin Surface Transducer	Acuson Antares	CH4-1, CH6-2, CW2, PH4-1	
	Endocavity Transducer		EC9-4	
	Skin Surface Transducer	Acuson CV70	3V2c, 7V3c, CW2, CW5	
		Acuson Cypress	3V2c, 4C1, 7V3c, Aux CW	
		Acuson S1000	4V1c, 4C1, 4V1, 6C2, 8V3, CW2, CW5	
	Endocavity Transducer		EC9-4, EV-8C4	
	Skin Surface Transducer	Acuson S2000	4V1c, 4C1, 4V1, 4V1c, 6C2, 8V3, CW2, CW5	
	Endocavity Transducer		EC9-4, EV-8C4	
	Ultrasound Transducer	Acuson SC2000 Prime	4V1c, 8V3, AUX CW, CW2	
		Acuson S3000	4C1, 4V1, 4V1c, 6C2, 8V3, CW2, CW5, EC9-4, EV-8C4	
Acuson SC2000		4V1c, 8V3, CW2		

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Siemens*	Ultrasound Transducer	Acuson Sequoia	EV-8C4
	Skin Surface Transducer		15L8, 3V2C, 4C1, 4V1/4V1c, 5V2c, 6C2, 6L3, 7V3c, 8C4, 8L5, 8L5T, 8V3c, 8V5, Aux CW
	Endocavity Transducer		EC-10C5
	Ultrasound Transducer	Acuson X700	4C1, 4V1c, 6C2, CW2, CW5, P8-4, VF16-5
	Skin Surface Transducer	Acuson X150	EV9-4
			CH5-2, P4-2, P8-4
			EC9-4
	Ultrasound Transducer	Acuson X300	CH5-2, CW2, CW5, EC9-4, EV9-4, P4-2, P8-4
		Acuson X300PE	CH5-2, CW2, CW5, EC9-4, EV9-4, P4-2, P8-4
		Acuson X500	CW2, CW5, EC9-4
		Sonoline G20	EC9-4
		Sonoline G40	CH5-2, EC9-4, EV9-4, P4-2
		Sonoline G50/G60	CW2, CW5, EC9-4
	TOE Transducer	Acuson Cypress	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson S1000	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson S2000	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson S3000	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson SC2000	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson SC2000 Prime	V5Ms (- SN 327XXXXX)
		Acuson SC2000 Prime	V5Ms (SN 328XXXXX -)
Acuson Sequoia		V5Ms (SN 327xxxxx & 328xxxxx)	
Acuson X300PE		V5Ms (SN 327xxxxx & 328xxxxx)	
Acuson X700		V5Ms (SN 327xxxxx & 328xxxxx)	
Acuson SC2000		Z6Ms	
SonoScape*	Ultrasound Transducer	E30 Series (E30/ E30 Pro/ E30 Exp/ E30S/ E30 Plus/ E30VO/ E29)	VC2-9A, VC2-9, VE3-10, VC6-2, VC8-3, VC311, VE9-5, VL12-5, VC6-2, EC2, EC9-5, 5V1, 6V1, 6V3, 6V4, 6V5, BCC9-4, BCC9-5, BCL10-5, 6V-A, 6V1A, 6V2A, 6V3A, 6V7, 8V-A, 6CT4, C3-10V, 2P1, 5P1, 10L1, C311, C321, C343, C344, C351, C352, C361, C362, C511, C521, C541, C542, C543, C611, C612, C721, L541, L561V, L662V, L741, L741V, L742, L743, L745, L746, L751, L752, L761V, L744, C312, C322, C345, C353, C354, C544, C613, C5-1, 3C-A, C1-6, 2P2, 3P1, 3P-A, 4P-A, 4P-B, 5P2, 7P-A, 7P-B, 8P1, 12P-A, 12P-B, P5-1, 3C-B, 9L-A, 13L-A, 18L-A, S1-5, C1-6A, C2-9, 12L-A, 12L-B, ML3-18, C9-5

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Sonosite Fujifilm	Portable Ultrasound System	FC1-System	Powerplug
		M-Turbo-System	M-Turbo-System
			M-Turbo-TTC
			Foot Paddle
			Powerplug
			Mini-Dock
		NanoMaxx-System	NanoMaxx-System
			Dock
			Powerplug
		M Turbo M-OB/GYN Office	(System & Dock)
No Series Specified	Micro Maxx		
No Series Specified	TITAN		
Sonosite Fujifilm	Ultrasound Transducer	No Series Specified	SLT
	Skin Surface Transducer	No Series Specified	P11x
			L52/L52e - Vet Only
	Ultrasound Transducer	Edge-System	P11x2
		Edge II-System	P11x2
		S-System	P11x2
		SII-System	P11x2
		M-Turbo-System	P11x2
		No Series Specified	D2/D2x
	Ultrasound System	iViz System	iViz System
			Protective Cover
	Ultrasound Unit	Edge-System	Edge-TTC
			Powerplug
			Mini-Dock
		Edge II-System	Edge II-System
			Edge-TTC
			Powerplug
			Mini-Dock
		S-System	S-System
			S-Series-TTC
Powerplug			
SII-System		SII-System	
		Powerplug	
Ultrasound Unit System	X-Porte-System	Powerplug	
TOE Transducer	No Series Specified	TEExi Transducer	

TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Canon (Toshiba)	Ultrasound Transducer	Aplio & Viamo Series	PLT-308P, PVT-770RT, PLT-705BTF/705BTH, PST25ST/50BT, PVT712BT, PST-25BT, PST-30BT, PLT-704SBT, , PLT704ST, PLT-1204ST, PVT-375ST
	Pencil Transducer		PC-20M, PC-50M
	TOE Transducer		PET-508MA, PET-512MA/512MC, PET511BTM
	Endocavity Transducer		PVT-661VT, PVT-781VT, PVT-781VTE, PVL-715RST, PVT-770RT, PVT-781VT, PVT-681MV
	Skin Surface Transducer		PVT-674BT, PLT-705BTH, PVT375ST *, PVT-7350BTP, PLT-1204BX, PLT-704SBT/705BT/1005BT, PLT704ST, PLT1204ST, PVT375ST, PST-25SX, PST-65AT, PLT1204ST, PLT-1204MV, PVT-375MV/382MV/675MV/675MVL, PST25ST, PST50BT, PST-30BT, PVT-375BT, PVT-375SC, PVT-382BT, PVT712BT, PLT-604AT, PLT-704AT, PLT-805AT, PLT-1204BT, PLT-1204BX, PLT-705BT, PLT-1005BT, PVT-647BT, PST-25ST/50BT, PVT-712BT, PST-25BT/50AT/65AT
	Intraoperative Transducer		PVT-745BTF, PVT-745BTH, PLT-705BTF, PVT-745BTV, PVT-745BTF/745BTH, PVT-745BTV, PLT-1202S
	Linear Transducer		PLT-308P
	Intraoperative Transducer		PVF-738F/738H/745V
	Nemio & Famio Series	Skin Surface Transducer	PLG-506M/805S
		Endocavity Transducer	PVG-366M/381M/681S/720S, PLQ-703A/805A/1203A, PVQ-375A/381A, PVQ-662A/674A
		Skin Surface Transducer	PVG-600S/601V/630V, PVQ-641V
		Endocavity Transducer	PLM-503AT/703AT/805AT
		Skin Surface Transducer	PLM-1204AT
		Intraoperative Transducer	PLM-1202S
		Skin Surface Transducer	PVF-381MT, PVF-575AT
		Endocavity Transducer	PVF-620ST
		Ultrasound Transducer	PVM-375AR
		Skin Surface Transducer	PVM-381AT
		Endocavity Transducer	PVM-620ST, PVM-621VT
		Skin Surface Transducer	PVM-651VT
		Skin Surface Transducer	PVM-656AT, PVM-662AT, PVM-672AT
		Ultrasound Transducer	PSM-20CT/25AT/30BT
		Intraoperative Transducer	PVM-656A
		Endocavity Transducer	PVF-738F*, PVF-738H*, PVF-745V*, PLG-308P, PLM-1202S
		Laparoscopic Transducer	PVG-600S*, PVG-601V*, PVG-630V*, PVM-651VT, PVM-740RT, PVL-715R, PVL-715RT, PVQ-641V, PVF-3620ST
		Skin Surface Transducer	PVM-787LA
	Pencil Transducer	SMA-736SA, PC-50V, PVG-366M*, PVG-381M*, PVG-681S*, PLF-308P, PLM-1204AT, PVF-381MT, PSM-37AT, PSM-70AT, PVF-575AT, PVQ-375A, PVQ-674A, PVM-375AT, PSM-37CT, PSM-50AT, PLG-506M*, PLG-805S*, PLQ-703A, PLQ-805A, PLQ-1203A, PLM-503AT, PLM-703AT, PLM-805AT, PSM-20CT, PSM-30BT	
		PC-19M	

## TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Canon (Toshiba)	Skin Surface Transducer	Xario Series	PSU-25BT/30BT & PVU-375BT/382BT
			PSU-50BT & PVU712BT
			PLU-704BT
			PVU-745BTV
			PVU-745BTF/745BTH/770ST
	Endocavity Transducer		PVL-715RS
	Skin Surface Transducer		PVU-674BT & PLU-1005BT, PVU-781VT/ 781VTE
	Endocavity Transducer		PVU-781VT
	Skin Surface Transducer		PLU-1204BT
	Skin Surface Transducer		PVU-745BTV, PVU-745BTF, PVU-745BTH, PVU-681MVL, PVU-375BT, PVU-382BT, PSU-50BT, PSU-25BT, PSU-30BT, PSU-70BT, PVU712BT, PLU-704BT
Endocavity Transducer	PVU-770ST		
TOE Probe	Aplio & Viamo Series	PET-508MA, PET-512MA, PET-512MC, PET511BTM	

**\* Compatible with the Tristel Sporicidal Wipe, part of the Tristel Trio Wipe System**

**Disclaimer:** Material compatibility testing performed via different methods including wiping and soaking. Tested materials subjected to varying number of cycles.

For more information contact Tristel team on [cooperation@tristel.com](mailto:cooperation@tristel.com)



**Created by:** Tristel Solutions Limited, Lynx Business Park, Cambs, UK, CB8 7NY  
T +44 (0) 1638 721500 - E [mail@tristel.com](mailto:mail@tristel.com) - W [www.tristel.com](http://www.tristel.com)

**Ireland:** Brennan & Company, 61 Birch Avenue, Stillorgan Industrial Park, Stillorgan, Co. Dublin, A94 XW68 - T 01 2952501 - E [enquiries@brennanco.ie](mailto:enquiries@brennanco.ie)

**Australian Sponsor:** Tristel Pty Ltd, 40/328 Reserve Road, Cheltenham, VIC 3192  
T 1300 680 898 - F +61 (0)3 9533 6193 - E [mail-au@tristel.com](mailto:mail-au@tristel.com)

For Tristel patent information please visit: <http://www.our-patents.info/tristel>

**New Zealand:** Tristel New Zealand Limited, 23 Birch Avenue, Judea, Tauranga P.O. Box 3110  
T +64 (0)7 5771560 - F +64 (0)7 5771567 - E [info@tristel.co.nz](mailto:info@tristel.co.nz)

**Hong Kong & Taiwan:** 21st Floor, 168 Electric Road, Fortress Hill, Hong Kong  
T +852 2895 6968 - F +852 2869 4388 - E [customerservice@tristel.com.hk](mailto:customerservice@tristel.com.hk)

**United Arab Emirates (UAE), Saudi Arabia, Qatar, Bahrain, Oman, Kuwait, Lebanon, Egypt, Iran, Iraq, Syria, Sudan, Yemen:** Fal Care FZE, Saif desk Z2-04, PO Box 514484, Sharja, United Arab Emirates - T 00971 54 707 049 - E [info@falcare.com](mailto:info@falcare.com)

**Israel:** Sachar Medical Technologies, 20 Zvi Berkman St., Petach Tiqwa, 4927976, Israel  
T 00972 544513925 - E [sachar.medtech@gmail.com](mailto:sachar.medtech@gmail.com)

Copyright © Tristel Solutions  
Mkt-Val-449-4  
December 2018