Tristel Duo™

OPH

High Level Disinfectant



Tristel Duo OPH is CE marked as a Class IIb Medical Device. It is approved as an Instrument Grade – High Level Disinfectant by the Australian Government Department of Health Therapeutic Goods Administration.

INTENDED USE

Tristel Duo OPH is designed specifically for the high level disinfection of instruments used in ophthalmology such as diagnostic lenses, tonometer prisms, slit lamps, pachymeters, A-scan and B-scan biometry probes.

THE TRISTEL CHEMISTRY

Tristel Duo OPH is based on Tristel's proprietary chlorine dioxide (ClO_2) chemistry, a well-documented and highly effective biocide. Tristel Duo OPH incorporates two separate compartments that contain Tristel Base solution (citric acid) and Tristel Activator solution (sodium chlorite). When mixed upon pressing the foam pump, Tristel's proprietary chlorine dioxide chemistry is generated. Tristel Duo OPH contains a non-ionic surfactant for added cleaning performance.

BIOCIDAL PERFORMANCE

Tristel Duo OPH is efficacious against bacterial spores, mycobacteria, viruses, fungi and vegetative bacteria. The product has been tested according to the relevant European Standards required for disinfectants used in medical areas. Duo OPH is effective against microorganisms of concern such as:

- *Clostridium sporogenes*
- *Mycobacterium terrae* (surrogate for *Mycobacterium tubercuolosis*)
- *Mycobacterium avium*
- Poliovirus Type 1
- Adenovirus Type 5
- Herpes Simplex Virus Type 1
- Vaccinia virus
- Norovirus
- *Aspergillus brasiliensis*
- *Candida albicans*
- Carbapenem-resistant Enterobacteriaceae (CRE) *Klebsiella pneumoniae*
- Vancomycin-resistant Enterococci (VRE) *Enterococcus faecium*
- *Pseudomonas aeruginosa*
- *Staphylococcus aureus*
- *Enterococcus hirae*

USER INSTRUCTIONS

- Do not use if the bottle is damaged.
- Disinfect hands and wear appropriate Personal Protective Equipment (PPE) when handling disinfectants and medical devices.
- If Tristel Duo OPH is used for the first time, depress the pump two to four times to prime it.
- If the surface is visibly soiled, follow the hospital protocol for cleaning instructions.
- For professional use.

Step 1
Dispense two aliquots of Tristel Duo OPH onto a dry wipe (Duo Wipes are recommended) or directly onto the instrument. Use the wipe to spread the foam over the surface of the medical device and ensure all areas are covered. Continuous wiping of the surface is not necessary and should be avoided to prevent shedding of the wipe.

Note: *Discard the used wipe in accordance with local regulations. Do not macerate. Do not re-use.*

Step 2
Leave the surface to dry to ensure a minimum contact time of two minutes.

Step 3
Rinse the medical device with water of appropriate quality.

Note: *Store the medical device in accordance with hospital protocols to prevent damage or recontamination.*

Warning

Base
Causes serious eye irritation.

Precautionary statements

Base
Wear eye protection.

IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

If eye irritation persists:
Get medical advice.

Manufactured by: Tristel Solutions Ltd

For Tristel patent information please visit **www.our-patents.info/tristel**

Contact Tristel, your local distributor or visit **www.tristel.com** for supporting documents such as safety data sheets, microbiological test data and reports.

Tristel™

WE HAVE CHEMISTRY.

Tristel Solutions Ltd, Lynx Business Park, Fordham Road, Snailwell, CB8 7NY, UK
+44 (0)1638 721500 · mail@tristel.com · www.tristel.com

Australian sponsor: Tristel Pty Ltd, 40/328 Reserve Road, Cheltenham, VIC 3192, AU
1300 680 898 · mail-au@tristel.com · www.tristel.com

New Zealand: Tristel New Zealand Ltd, 23 Birch Avenue, Tauranga 3110, NZ
+64 (0)7 5771560 · info@tristel.co.nz · www.tristel.com

EC REP Tristel NV, Smallandlaan 14 B,
2660 Antwerp, Belgium

TR/DOO/601 · 18.03.2020 · V4