

NAVIGATING THE SHADOWS

– Understanding 6 Uncertainties with UV-C Technologies

Introduction

Machines offering ultraviolet-C (UV-C) light for the high-level disinfection of medical devices are rapidly emerging worldwide. This surge in the use of UV-C requires healthcare practitioners to take extra precautions before embedding unconventional technology into standard procedures.

What is UV-C?

Ultraviolet-C radiation is a wavelength of light that falls between the spectrum of X-rays and visible light. Light within this range has been shown to inactivate certain microorganisms through a specific reaction pathway. This is typically achieved using low-pressure mercury lamps. However, recently, LED lamps have been utilised as an alternative.



Ultraviolet (UV) light destroys microorganisms by targeting the DNA or RNA within microorganisms. This causes a photochemical reaction, creating abnormal links within the DNA or RNA. The chemical structure is distorted which interferes with the normal function of the microorganism, leading to mutations or inactivation.


How is it used?

UV light is used in various settings, including water treatment systems and piping networks, as well as healthcare and food industry settings, including the disinfection of ventilation systems, surfaces, and devices. Its use in healthcare is becoming increasingly common, particularly for the disinfection of non-lumened, semi-critical medical devices.

A typical UV-C machine comprises a hard plastic cabinet housing multiple UV lamps within a mirrored glass chamber. A portion of the medical device is hung inside the chamber, with the door closed to protect the end user, and UV-C light is emitted onto the device.

This article outlines six critical considerations before adopting UV-C technology for your hospital medical device decontamination procedure.

The Spectrum of Light



1. Lack of Standardisation and Data

EN 14885 *Chemical disinfectants and antiseptics* – Application of European Standards for chemical disinfectants and antiseptics plays a crucial role in ensuring the efficacy, safety, and quality of chemical disinfectants and antiseptics used in various settings, including healthcare. Chemical disinfectants in Europe and the UK follow this guidance as it specifies the requirements for the testing of chemical disinfectants and antiseptics to support the antimicrobial claims they make.

(Bolten, Schmidt and Steinhauer, 2022)¹

The standardisation of disinfectant test methods is critical because it:

- enables manufacturers of chemical disinfectant products to choose the appropriate laboratory methods to support efficacy claims,
- enables end users of chemical disinfectant products to assess the information provided by the manufacturer and
- supports regulatory authorities in assessing claims made by manufacturers of disinfectant products.

No standardised method for testing a UV-C machine intended for the disinfection of semi-critical devices has been published. Manufacturers of UV-C machines are using the standard methods within the **EN 14885** to demonstrate the effectiveness of their systems. However, these standards were developed specifically for chemical disinfection systems. They include practical methods that simulate how chemical products are intended to be used in a real-world application and include microorganisms that are explicitly challenging for chemical disinfectants.

UV-C machines are not chemical systems. Existing test methods within EN 14885 are not designed for non-chemical products, which inevitably leads to the UV-C manufacturers modifying these methods to suit their products. When deviations from the standard methods are made, the benefits of standardisation are lost, and we are left with no assurance that the product has been challenged appropriately.

The **EN 14885**² states:

'In certain cases, it can be necessary or recommendable to modify even the test organism(s) to match the requirements of the area.'

Conformity to the standard used shall not be claimed, but it should be stated that the product was tested in accordance with the principles of the standard.'

*'Where there is no appropriate standard for an application within a specific area, a standard from another area may be recommended for use. **If later on an appropriate standard is published, this new standard shall be used.'***

The **BS 8628** is the first of its kind and provides methods for quantitative testing of automated UV disinfection activities by direct illumination. This standard is now the most appropriate standard for testing UV machines intended for disinfection and is derived from the pre-existing standard **EN 17272**, which outlines methodologies for the microbiological efficacy testing of airborne disinfection processes/systems.

In this case, **BS 8628** is the minimum we should expect from all UV-C machines claiming to high-level disinfect semi-critical medical devices on the market.

2. Compatibility

Plastic polymers make up a significant proportion of modern semi-critical devices. This includes ultrasound devices and endoscopes, which typically contain plastics such as Acrylonitrile Butadiene Styrene (ABS), Polypropylene (PP), Polycarbonate (PC), and Polyvinyl Chloride (PVC). These plastics offer a balance of properties such as mechanical strength, chemical resistance, flexibility, transparency, and ease of processing, making them suitable for various components and applications within ultrasound devices.

UV-C light has been shown to cause damage to plastics through consistent exposure. A recent study assessed the impact of UV-C light exposure on 10 plastic materials commonly used in healthcare. All materials showed degradation, however some **plastics such as ABS and Polycarbonate showed evidence of significant surface damage.**³ These materials are commonly used for device surroundings and casings.

It was concluded that: ***'exposure to UV-C should be limited for these types of materials as much as possible in healthcare facilities, especially where UV-C devices are routinely used'***.

(Teska et al., 2020)³

3. Requirement for Manual Cleaning

Cleaning is regarded as the most critical step in the decontamination procedure.

*'Thorough cleaning is required before high-level disinfection and sterilization because **inorganic and organic materials that remain on the surfaces of instruments interfere with the effectiveness of these processes.**'*

(Rutala and Weber, 2016)⁴

Ultimately, if the cleaning process is compromised, it is likely that the subsequent disinfection process will be adversely affected. Cleaning holds particular importance for UV-C systems, where no manual abrasion is used. If cleaning fails, soiling will persist even after exposure to UV-C disinfection, rendering the entire process ineffective. In contrast, manual processes like wiping, involves an additional abrasion step, facilitating the removal of any remaining soiling.




It is crucial to emphasise that cleaning agents must be thoroughly removed from the device prior to disinfection with UV-C, and the device should be completely dry before being placed in a UV chamber. Any residual surface moisture can hinder the delivery of the required UV dose to the surface.

4. Shadowing

UV light travels in a straight line, and objects obstructing its path can create shadows by preventing the passage of light. When UV light is blocked, it fails to reach specific areas during disinfection, posing a risk of inadequate disinfection in those regions.



Some ultrasound devices may have complex shapes or surface textures that create a challenge in achieving uniform exposure to UV light, potentially resulting in areas where disinfection will be unsuccessful.



Surface topography refers to the physical features of a surface/device and includes complexities such as indentations, contours, textures, and any irregularities in shape. These complexities can block the light from reaching the surface beneath, causing shadowing. Most chemical products will not have an issue reaching these areas as additional attention can be paid by the user. Products such as foams may even get into surface scratches where pathogens reside.

In contrast to alternative disinfection methods, UV-C disinfection offers

limited penetration depth. Consequently, it cannot effectively penetrate solid or liquid residues that may persist on surfaces with equal intensity. Therefore, if any soiling or moisture remains on a surface prior to UV-C exposure, the light will not reach all areas for disinfection, potentially leaving spots where microorganisms could persist.

Example:

UV disinfection is commonly employed for disinfecting organic fruit produce. A study evaluated how the surface roughness of fruit impacts the effectiveness of UV-C treatment.

Fruits like apples, which have smoother surfaces, were shown to have significantly higher reductions in organisms compared to those with more intricate surfaces, such as peaches, raspberries, and even pears. It was concluded that this was down to the subtle roughness of the fruit surface compared to that of an apple.



The premise of this study applies to all surface types. Ultimately, surfaces with varying textures or complexities

are anticipated to pose a challenge for light exposure, making it difficult for light to reach all areas of the surface.

(M.S. Roopesh et al., 2012)⁵

Surface scratches on older, used devices can equally contribute to the effect of shadowing. UV-C radiation is capable of damaging plastics integral to ultrasound devices, and this raises concerns about potential surface cracking with prolonged exposure. Such cracks could exacerbate the risk of shadowing, potentially shielding harmful pathogens from disinfection.

5. Other Areas of the Probe and Global Guidance

UV systems typically do not address the entire device, including the **probe cable, plug, holder, and machine**. Published literature has shown that these areas can become readily contaminated with pathogenic microorganisms.

'Probe cords and machine keyboards present significant sources of infection, and this can include potential pathogens.'

(Westerway et al., 2017)⁶



UV systems are designed solely to disinfect the portion of the probe that come into contact with the patient. This does not consider the possibility that staff performing the procedure may come into contact with other areas of the device during a patient procedure, inadvertently causing cross-contamination. These areas should be considered equal to the insertable portion of the device and be disinfected to the same level, which is not possible with most UV systems.

Guidance:

World Federation for Ultrasound in Medicine and Biology (WFUMB) – Guidelines for Cleaning Transvaginal Ultrasound Transducers Between Patients, 2017 states:

'An additional consideration is the fact that the transducer handle and cable can also become contaminated and may also require disinfection.'

(Abramowicz et al., 2017)⁷

Also, the **Infection Prevention and Control in Ultrasound. Best Practice Recommendations from the European Society of Ultrasound Working Group, 2017** states:

'Thorough decontamination of US transducers and any equipment in direct patient contact before and after every patient, to the level required for specific procedures and in compliance with manufacturer specifications to avoid transducer surface damage, should be carried out.

This includes regular decontamination of US keyboard/console and any cables.'

'All US equipment in direct or indirect patient contact must be thoroughly cleaned and disinfected at the start of the examination and after every patient. This includes the US transducer with handle, cable and transducer holder (as far as possible) as well as additional devices which may be used during diagnostic or interventional procedures such as US fusions sensors/cables...

Contamination of US equipment may be underestimated.'

(Nyhsen et al., 2017)⁸



6. Repair Mechanisms

Research has indicated that following UV-C exposure, microorganisms possess two mechanisms for self-repair, referred to as **Photoreactivation** and **Dark Repair**.

Photoreactivation:

Photoreactivation is a biological process observed in certain organisms, particularly bacteria and viruses. Exposure to visible light following damage from UV triggers enzymatic repair mechanisms within the organism's DNA.

When microorganisms are exposed to UV light, particularly in the UV-C range, the UV photons cause damage to the DNA structure, resulting in the formation of pyrimidine dimers. However, in the presence of visible light, specific enzymes called photolyases become activated. These photolyases can recognize and bind to the damaged DNA sites, facilitating the reversal of the DNA damage. Essentially, **photoreactivation allows microorganisms to repair the DNA damage caused by UV radiation light**, increasing their chances of survival.

Dark Repair:

Dark repair, also known as nucleotide excision repair, is another biological process observed, particularly in bacteria. It involves repairing DNA damage caused by UV radiation in the absence of visible light.

In the **absence of visible light**, specific enzymes and repair pathways become activated to identify and repair the damaged DNA segments. During dark repair, enzymes recognize the lesions in the DNA strand and excise the damaged portion, replacing it with newly synthesized DNA and restoring the original DNA sequence.

Dark repair is an essential mechanism for microorganisms to maintain genome integrity and ensure survival following exposure to UV.

These mechanisms can revive a considerable portion of the microorganisms originally killed by UV-C light, reducing the overall effectiveness of the disinfection process.

Example:

A study has shown that as much as **60%** of microorganisms originally killed with UV-C can be reversed.

(Song, Mohseni and Taghipour, 2019)⁹

This study was conducted with *E. coli* but many other bacteria contain enzymes and repair mechanisms that may allow for photoreactivation and photorecovery from UV exposure.

(Kowalski, 2009)¹⁰

Antimicrobial Resistance:

Repair mechanisms alter the DNA/RNA, which creates a risk of microorganism **mutations** during the repair process. These mutations could render microorganisms more pathogenic or resistant to future disinfection attempts. This could potentially diminish the effectiveness of UV-C further over time.

(Shibai et al., 2017)¹¹

As a result, it has been recommended that UV-C disinfection may not be entirely reliable as a stand-alone disinfection method and requires the implementation of complementary disinfection strategies to ensure thorough microbial eradication.

(Demeersseman et al., 2023)¹²

Conclusion

In conclusion, the widespread adoption of Ultraviolet-C (UV-C) technologies for high-level disinfection in healthcare settings has brought about significant advancements but has also raised several critical uncertainties and challenges that should be considered before the technology is adopted.

From the lack of standardised testing methods to issues regarding material compatibility, manual cleaning requirements, shadowing effects, addressing other areas of the device, and the concern of UV-C repair mechanisms, healthcare practitioners must navigate through these complexities before integrating UV-C technology into standard procedures.

Disclaimer

This document serves to provide valuable insights into available published literature and guides informed decision-making regarding the implementation of UV-C disinfection technologies in healthcare settings.

References

- ¹ Bolten, A., Schmidt, V. and Steinhauer, K. (2022). Use of the European standardization framework established by CEN/TC 216 for effective disinfection strategies in human medicine, veterinary medicine, food hygiene, industry, and domestic and institutional use - a review. *DOAJ (DOAJ: Directory of Open Access Journals)*, 17, pp.Doc14–Doc14.
- ² British Standards Institution (BSI). EN 14885: Chemical Disinfectants and Antiseptics - Application of European Standards for Chemical Disinfectants and Antiseptics. Brussels: European Committee for Standardisation (CEN); 2022.
- ³ Teska, P., Dayton, R., Li, X., Lamb, J. and Strader, P. (2020). Damage to Common Healthcare Polymer Surfaces from UV Exposure. *Nano LIFE*, 10(03), p.2050001.
- ⁴ Rutala, W. and Weber, D. (2008) Healthcare Infection Control Practices Advisory Committee, Centers for Disease Control and Prevention (CDC). Guideline for Disinfection and Sterilization in Healthcare facilities, 2008. Centers for Disease Control and Prevention.
- ⁵ M.S. Roopesh, Lu, X., Sablani, S.S., Sunil Kumar Insan, Adhikari, A., Killinger, K., Rasco, B., Dhingra, A., Bandyopadhyay, A. and Annapure, U.S. (2012). Inactivation of *Escherichia coli* Population on Fruit Surfaces Using Ultraviolet-C light: Influence of Fruit Surface Characteristics. *Food and Bioprocess Technology*, 6(11), pp.2959–2973.
- ⁶ Westerway, S.C., Basseal, J.M., Brockway, A., Hyett, J.A. and Carter, D.A. (2017). Potential Infection Control Risks Associated with Ultrasound Equipment – A Bacterial Perspective. *Ultrasound in Medicine & Biology*, 43(2), pp.421–426.
- ⁷ Abramowicz, J.S., Evans, D.H., Fowlkes, J.B., Maršal, K. and terHaar, G. (2017). Guidelines for Cleaning Transvaginal Ultrasound Transducers Between Patients. *Ultrasound in Medicine & Biology*, 43(5), pp.1076–1079.
- ⁸ Nyhsen, C.M., Humphreys, H., Koerner, R.J., Grenier, N., Brady, A., Sidhu, P., Nicolau, C., Mostbeck, G., D'Onofrio, M., Gangi, A. and Claudon, M. (2017). Infection prevention and control in ultrasound - best practice recommendations from the European Society of Radiology Ultrasound Working Group. *Insights into Imaging*, 8(6), pp.523–535.
- ⁹ Song, K., Mohseni, M. and Taghipour, F. (2019a). Mechanisms investigation on bacterial inactivation through combinations of UV wavelengths. *Water Research*, 163, p.114875.
- ¹⁰ Kowalski, W. (2009). Ultraviolet Germicidal Irradiation Handbook.
- ¹¹ Shibai, A., Takahashi, Y., Ishizawa, Y., Motooka, D., Nakamura, S., Ying, B.-W. and Tsuru, S. (2017). Mutation accumulation under UV radiation in *Escherichia coli*. *Scientific Reports*, 7(14531).
- ¹² Demeersseman, N., Saegeman, V., Cossey, V., Devriese, H. and Schuermans, A. (2023). Shedding a light on ultraviolet-C technologies in the hospital environment. *Journal of Hospital Infection*, 132, pp.85–92.